



EDWARD M. BIRN  
Director (Direktot)  
ELIZABETH T. FISHER  
Deputy Director (Sigundo Direktot)

**DEPARTMENT OF  
ADMINISTRATION**  
DIPATTAMENTON ATMENESTRASION

DIRECTOR'S OFFICE  
(Ufisinan Direktot)  
Telephone (Telifon): (671) 475-1101/1250



LOURDES A. LEON GUERRERO  
Governor (Maga'håga)  
JOSHUA F. TENORIO  
Lt. Governor (Sigundo Maga'låhi)

July 27, 2023

Invitation For Bid  
GSA-051-23

**AMENDMENT #5**  
**Fourier Transform Infrared FTIR Spectrometer**

1.) Amend and replace page 38 of 38 with attached "Revised page 38 of 38" dated 07/27/2023

All others remain unchanged.

JoyJean Arceo  
Acting Chief Procurement Officer\*

\*By appointment of Director of the Department of Administration effective July 26, 2023 through September 26, 2023.

Please Print  
Acknowledgement Copy (Re-Fax to GSA)

Received By: \_\_\_\_\_  
Date: \_\_\_\_\_  
Company Name: \_\_\_\_\_  
Fax to: 475-1727 or 472-4217  
Email to: [gsaprocurement@gsa.gov](mailto:gsaprocurement@gsa.gov)

- ❖ Dimensions: 1.024 x 3.94 x 14.49 in  
(26 x 10 x 36.8cm) \_\_\_\_\_
  - ❖ Weight: 27.56 lbs (12.5 kg) \_\_\_\_\_
  - ❖ Power cords are within North American Standards \_\_\_\_\_
  - ❖ Approvals: ENERGY STAR V2.0 (USA), FCC Part 15 \_\_\_\_\_
  - ❖ Class B, NOM \_\_\_\_\_
  - ❖ TUV C-US certified \_\_\_\_\_
- Installation, Warranty and Service:
    - ❖ Parts-Only Warranty two (2) years \_\_\_\_\_
    - ❖ Must Extend factory parts warranty on instrument for one (1) year \_\_\_\_\_
    - ❖ Must Extend warranty up to five (5) years, with first year included with system. \_\_\_\_\_
    - ❖ Onsite Installation, commissioning and training must be performed manufacturer's trained service engineer and applications chemist. Certificates of training must accompany all offers. \_\_\_\_\_
    - ❖ Service must be available with capability for after-sales support and remote diagnostics via telephone and/or email. \_\_\_\_\_

NOTE:

The manufacturer must possess ISO 9001 certification for the manufacturing facility where the spectrometer is produced. Installation shall include familiarization with the instrument and all its components to be conducted by a trained/qualified Engineer. Training of the end-users must at the least, but not limited to, Basic Theory, Sample handling techniques, hands-on training, Software and interpretations of results, and maintenance and troubleshooting. Programming of the instrument and all its related components must be tested and operational. Total cost shall be inclusive of all labor, cables, data wire, connectors, materials, equipment, tools, supervision, overhead, profit, end-user training, cleaning and disposal (if necessary), and GRT to complete the installation.

**Bidding On**  
**Manufacturer:** \_\_\_\_\_  
**Model:** \_\_\_\_\_  
**Place of Origin:** \_\_\_\_\_  
**Date of Delivery:** \_\_\_\_\_

**Funded by: 2020 Paul Coverdell Grant**

These specifications were drafted by Ms. Analyn Akigami, Criminalist III, and Ms. Monica Salas, Chief Criminalist and approved by: Mr. Stephen Ignacio, Chief , Guam Police Department.